

## Deutsche Akkreditierungsstelle GmbH German Accreditation Body

**Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV**

Signatory to the Multilateral Agreements of  
EA, ILAC and IAF for Mutual Recognition

# Accreditation



The Deutsche Akkreditierungsstelle GmbH (German Accreditation Body) attests that the calibration laboratory

**MAGNET-PHYSIK Dr. Steingroever GmbH**  
**Emil-Hoffmann-Straße 3, 50996 Köln**

is competent under the terms of DIN EN ISO/IEC 17025:2005 to carry out calibrations in the following fields:

**Magnetic quantities**

- **Magnetic flux density**
- **Magnetic field strength**

The accreditation certificate shall only apply in connection with the notice of accreditation of 23.06.2014 with the accreditation number D-K-19647-01 and is valid until 22.06.2019. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 04 pages.

Registration number of the certificate: **D-K-19647-01-00**

# Deutsche Akkreditierungsstelle GmbH

Office Berlin  
Spittelmarkt 10  
10117 Berlin

Office Frankfurt am Main  
Gartenstraße 6  
60594 Frankfurt am Main

Office Braunschweig  
Bundesallee 100  
38116 Braunschweig

The publication of extracts of the accreditation certificate is subject to the prior written approval by Deutsche Akkreditierungsstelle GmbH (DAkkS). Exempted is the unchanged form of separate disseminations of the cover sheet by the conformity assessment body mentioned overleaf.

No impression shall be made that the accreditation also extends to fields beyond the scope of accreditation attested by DAkkS.

The accreditation was granted pursuant to the Act on the Accreditation Body (AkkStelleG) of 31 July 2009 (Federal Law Gazette I p. 2625) and the Regulation (EC) No 765/2008 of the European Parliament and of the Council of 9 July 2008 setting out the requirements for accreditation and market surveillance relating to the marketing of products (Official Journal of the European Union L 218 of 9 July 2008, p. 30). DAkkS is a signatory to the Multilateral Agreements for Mutual Recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Cooperation (ILAC). The signatories to these agreements recognise each other's accreditations.

The up-to-date state of membership can be retrieved from the following websites:

EA: [www.european-accreditation.org](http://www.european-accreditation.org)

ILAC: [www.ilac.org](http://www.ilac.org)

IAF: [www.iaf.nu](http://www.iaf.nu)

# Deutsche Akkreditierungsstelle GmbH

## Annex to the Accreditation Certificate D-K-19647-01-00 according to ISO/IEC 17025:2005

Period of validity: 23.06.2014 to 22.06.2019

Date of issue: 23.06.2014

Holder of certificate:

**MAGNET-PHYSIK Dr. Steingroever GmbH**  
**Emil-Hoffmann-Straße 3, 50996 Köln**

Head: Dr. rer. nat. Klaus Wagner  
Deputy: B.Sc. Jan Kai Hutzenlaub

Accredited as calibration laboratory since: 05.01.2001

Calibrations in the fields:

**Magnetic quantities**

- **Magnetic flux density**
- **Magnetic field strength**

Abbreviations used: see last page

Annex to the accreditation certificate D-K-19647-01-00

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability <sup>1)</sup>	Remarks
Magnetic field strength Reference magnet systems Reference magnets	8 A/m to 64·10 <sup>3</sup> A/m  > 64·10 <sup>3</sup> A/m to 13·10 <sup>5</sup> A/m  > 13·10 <sup>5</sup> A/m to 24·10 <sup>5</sup> A/m	23 °C Fluxmeter and field sensing coil, magnetometer and Hall probe  Nuclear magnetic resonance magnetometer (air gap > 8.5 mm, high homogeneity), otherwise fluxmeter and field sensing coil or magnetometer and Hall probe  Fluxmeter and field sensing coil, magnetometer and Hall probe	1.1·10 <sup>-3</sup> + 0.24·(A/m)/H  0.14·10 <sup>-3</sup>  2·10 <sup>-3</sup>	H = measurement result
Magnetic field strength Hall effect and other magnetometers together with probes	0.8 A/m to 80 A/m > 80 A/m to 8·10 <sup>2</sup> A/m > 8·10 <sup>2</sup> A/m to 10.7·10 <sup>3</sup> A/m > 10.7·10 <sup>3</sup> A/m to 64·10 <sup>3</sup> A/m > 64·10 <sup>3</sup> A/m to 13·10 <sup>5</sup> A/m > 13·10 <sup>5</sup> A/m to 24·10 <sup>5</sup> A/m	DC; 23 °C Measurement in a field generating system  a) Substitution method b) By means of a calibrated reference magnet c) By means of a calibrated field generating coil	1·10 <sup>-3</sup> + 7.2·10 <sup>-3</sup> (A/m)/H 1·10 <sup>-3</sup> 1.6·10 <sup>-3</sup> 2.6·10 <sup>-3</sup> 0.17·10 <sup>-3</sup> 3·10 <sup>-3</sup>	H = measurement result
Magnetic flux density Reference magnet systems Reference magnets	1·10 <sup>-5</sup> T to 8·10 <sup>-2</sup> T  > 8·10 <sup>-2</sup> T to 1.7 T  > 1.7 T to 3 T	23 °C Fluxmeter and field sensing coil, magnetometer and Hall probe  Nuclear magnetic resonance magnetometer (air gap > 8.5 mm, high homogeneity), otherwise fluxmeter and field sensing coil or magnetometer and Hall probe  Fluxmeter and field sensing coil, magnetometer and Hall probe	1.1·10 <sup>-3</sup> + 0.3·10 <sup>-6</sup> T/B  0.14·10 <sup>-3</sup>  2·10 <sup>-3</sup>	B = measurement result
Magnetic flux density Hall effect and other magnetometers together with probes	1·10 <sup>-6</sup> T to 1·10 <sup>-4</sup> T > 1·10 <sup>-4</sup> T to 1·10 <sup>-3</sup> T > 1·10 <sup>-3</sup> T to 13.5·10 <sup>-3</sup> T > 13.5·10 <sup>-3</sup> T to 8·10 <sup>-2</sup> T > 8·10 <sup>-2</sup> T to 1.7 T > 1.7 T to 3 T	DC; 23 °C Measurement in a field generating system  a) Substitution method b) By means of a calibrated reference magnet c) By means of a calibrated field generating coil	1·10 <sup>-3</sup> + 9·10 <sup>-9</sup> T/B 1·10 <sup>-3</sup> 1.6·10 <sup>-3</sup> 2.6·10 <sup>-3</sup> 0.17·10 <sup>-3</sup> 3·10 <sup>-3</sup>	B = measurement result
Magnetic flux density Nuclear magnetic resonance magnetometers	8·10 <sup>-2</sup> T to 1.7 T	DC; 23 °C Measurement in a field generating system, substitution method	0.17·10 <sup>-3</sup>	Comparison of nuclear magnetic resonance magnetometers

<sup>1)</sup> The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-19647-01-00

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability <sup>1)</sup>	Remarks
Magnetic moment Reference magnet systems Magnet etalon	$8 \cdot 10^{-4} \text{ Am}^2$ to $24 \cdot 10^{-3} \text{ Am}^2$ > $24 \cdot 10^{-3} \text{ Am}^2$ to $54 \text{ Am}^2$	23 °C Fluxmeter and Helmholtz coils	$3 \cdot 10^{-3} + 12 \cdot 10^{-6} \text{ Am}^2/m$ $2.1 \cdot 10^{-3} + 33 \cdot 10^{-6} \text{ Am}^2/m$	$m$ = measurement result
Magnetic dipole moment Reference magnet systems Magnet etalon	$1 \cdot 10^{-9} \text{ Vsm}$ to $3 \cdot 10^{-8} \text{ Vsm}$ > $3 \cdot 10^{-8} \text{ Vsm}$ to $68 \cdot 10^{-6} \text{ Vsm}$	23 °C Fluxmeter and Helmholtz coils	$3 \cdot 10^{-3} + 15 \cdot 10^{-12} \text{ Vsm/j}$ $2.1 \cdot 10^{-3} + 42 \cdot 10^{-12} \text{ Vsm/j}$	$j$ = measurement result
Sensitivity Magnetic field sensors	$1 \cdot 10^{-2} \text{ V/T}$ to $1 \cdot 10^3 \text{ V/T}$	DC; 23 °C at $1 \cdot 10^{-3} \text{ T}$ to $13.5 \cdot 10^{-3} \text{ T}$ at > $13.5 \cdot 10^{-3} \text{ T}$ to $8 \cdot 10^{-2} \text{ T}$ at > $8 \cdot 10^{-2} \text{ T}$ to $1.7 \text{ T}$ at > $1.7 \text{ T}$ to $3 \text{ T}$ Measurement in a field generating system a) Substitution method b) By means of a calibrated reference magnet c) By means of a calibrated field generating coil	$1.6 \cdot 10^{-3} + 4 \cdot 10^{-6} \text{ V/(S \cdot B)}$ $2.6 \cdot 10^{-3} + 4 \cdot 10^{-6} \text{ V/(S \cdot B)}$ $0.2 \cdot 10^{-3} + 4 \cdot 10^{-6} \text{ V/(S \cdot B)}$ $3 \cdot 10^{-3} + 4 \cdot 10^{-6} \text{ V/(S \cdot B)}$	$S = U/B$ = sensitivity value at given supply voltage or supply current, $U$ = sensor output voltage, $B$ = magnetic flux density
Magnetic flux Fluxmeters	$1 \cdot 10^{-4} \text{ Vs}$ to $2.7 \cdot 10^{-4} \text{ Vs}$ > $2.7 \cdot 10^{-4} \text{ Vs}$ to $8 \cdot 10^{-4} \text{ Vs}$	DC; 23 °C Calibrated Helmholtz coil and field sensing coil	$4.1 \cdot 10^{-7} \text{ Vs}/\Phi + 1.2 \cdot 10^{-3}$ $8 \cdot 10^{-7} \text{ Vs}/\Phi + 2.2 \cdot 10^{-3}$	$\Phi$ = measurement result of magnetic flux
	> $8 \cdot 10^{-4} \text{ Vs}$ to $3.4 \cdot 10^{-2} \text{ Vs}$ > $3.4 \cdot 10^{-2} \text{ Vs}$ to $0.34 \text{ Vs}$	DC; 23 °C Electromagnet with yoke, measurement of flux density by means of nuclear magnetic resonance, calibrated field sensing coil	$1.5 \cdot 10^{-3}$ $3 \cdot 10^{-3}$	
	$1.98 \cdot 10^{-3} \text{ Vs}$	DC; 23 °C Reference magnet calibrated by nuclear magnetic resonance, calibrated field sensing coil	$1.6 \cdot 10^{-3}$	
Magnetic flux Flux etalon	$1 \cdot 10^{-2} \text{ Vs}$	23 °C Measurement by means of a calibrated fluxmeter	$2.2 \cdot 10^{-3}$	
Winding area Field sensing coils	$1 \cdot 10^{-4} \text{ m}^2$ to $20 \text{ m}^2$	DC; 23 °C a) Measurement of flux in a reference magnet or in a field generating coil b) Comparison with a known winding area	$1.5 \cdot 10^{-3} + 1.5 \cdot 10^{-6} \text{ m}^2/A$	$A$ = measurement result of winding area
Measuring constant Moment detection coils	$1 \cdot 10^{-6} \text{ m}$ to $1 \cdot 10^{-2} \text{ m}$ > $1 \cdot 10^{-2} \text{ m}$ to $1 \text{ m}$	DC; 23 °C Substitution method, comparison with a calibrated moment detection coil by means of a permanent magnet	$0.6 \cdot 10^{-3}$ $1 \cdot 10^{-3} + 4 \cdot 10^{-2} \text{ k}_M/m$	$k_M$ = measurement result

<sup>1)</sup> The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

**Annex to the accreditation certificate D-K-19647-01-00**

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability <sup>1)</sup>	Remarks
Magnetic field to current constant Field generating coils	$1 \text{ m}^{-1}$ to $1 \cdot 10^2 \text{ m}^{-1}$ $> 1 \cdot 10^2 \text{ m}^{-1}$ to $1 \cdot 10^6 \text{ m}^{-1}$	DC; 23 °C Measurement of the current through the coil and the magnetic field strength in the coil centre. Measurement of the current through the coil and the magnetic field strength along the coil axis.	$1 \cdot 10^{-3} + 4 \cdot 10^{-2} / (m \cdot k_F)$ $0.6 \cdot 10^{-3}$	$k_F = H_{Zentr} / I$ = value of the magnetic field to current constant $I$ = measurement result of current through the coil $H_{Zentr}$ = measure- ment result of magnetic field strength in the centre of the coil

**Abbreviations used:**

EA-4/02 „Expression of the Uncertainty of Measurement in Calibration“

<sup>1)</sup> The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.